



OIL ANALYSIS REPORT

WEAR	NORMAL
CONTAMINATION	NORMAL
FLUID CONDITION	NORMAL

Area
[531674]
Machine Id
1995103
Component
Unknown Component
Fluid
{not provided} (--- GAL)

RECOMMENDATION

This is a baseline read-out on the submitted sample.

Test	UOM	Method	Limit/Abn	Current	History1	History2
Sample Number		Client Info		PP	---	---
Sample Date		Client Info		17 Apr 2024	---	---
Machine Age	hrs	Client Info		0	---	---
Oil Age	hrs	Client Info		0	---	---
Filter Age	hrs	Client Info		0	---	---
Oil Changed		Client Info		N/A	---	---
Filter Changed		Client Info		N/A	---	---
Sample Status				NORMAL	---	---

WEAR

{not applicable}

White Metal	scalar	Visual*	NONE	NONE	---	---
Yellow Metal	scalar	Visual*	NONE	NONE	---	---

CONTAMINATION


{not applicable}

Water		WC Method		NEG	---	---
Silt	scalar	Visual*	NONE	NONE	---	---
Debris	scalar	Visual*	NONE	VLITE	---	---
Sand/Dirt	scalar	Visual*	NONE	NONE	---	---
Appearance	scalar	Visual*	NORML	NORML	---	---
Odor	scalar	Visual*	NORML	NORML	---	---

FLUID CONDITION

ASTM D92 - Open-Cup Flash Point = 232°C.

COC Flash Point	°C	ASTM D92*		232	---	---
-----------------	----	-----------	--	------------	-----	-----



Laboratory : WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9
Sample No. : PP **Received** : 19 Apr 2024
Lab Number : 02630423 **Tested** : 19 Apr 2024
Unique Number : 5763555 **Diagnosed** : 19 Apr 2024 - Kevin Marson
Test Package : TEST (Additional Tests: COC Flash)

To discuss this sample report, contact Customer Service at 1-800-268-2131.
 Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab.
 Validity of results and interpretation are based on the sample and information as supplied.

TESTMARK LABORATORIES
 6820 KITIMAT ROAD
 MISSISSAUGA, ON
 CA L5L 5T5
 Contact: Mississauga Sample
 mississauga.oils@testmark.ca
 T: (905)821-1112
 F: