


## Machine Id LIEBHERR R914 054895-1511

Component Coolant

HYBRID (HOAT) COOLANT (	GAL)						
RECOMMENDATION	Test	UOM	Method	Limit/Abn	Current	History1	History2
The fluid is suitable for further service. Resample at the next service interval to monitor.	Sample Number		Client Info		LH0289510		
	Sample Date		Client Info		21 May 2024		
	Machine Age	hrs	Client Info		870		
	Oil Age	hrs	Client Info		0		
	Oil Changed		Client Info		Not Changd		
	Sample Status				NORMAL		
CORROSION	Iron	ppm	ASTM D5185(m)	>15	<1		
All metal levels are normal indicating no corrosion in the cooling system.	Aluminum	ppm	ASTM D5185(m)		<1		
	Copper	ppm	ASTM D5185(m)		<1		
	Lead	ppm	ASTM D5185(m)		0		
	Tin	ppm	ASTM D5185(m)	>10	0		
	Silver	ppm	ASTM D5185(m)	>10	<1		
	Zinc	ppm	ASTM D5185(m)	>10	3		
CONTAMINANTS	Oalai:		ACTM DE105(**)	100			
	Calcium	ppm	ASTM D5185(m)		9		
There is no indication of any contamination in the coolant.	Magnesium	ppm	ASTM D5185(m) Visual*	>40	9 NODMI		
	Appearance Odor	scalar	Visual*	NORML	NORML NORML		
	Coolant Appearance	scalar	Visual*	Clear	Cloudy		
			visuai	Oleai	·····		
COOLANT CONDITION	Boiling Point	°C	WC Method*		108		
The pH level of this fluid is within the acceptable limits. The reserve alkalinity of this fluid is acceptable.	Specific Gravity		ASTM D1298*		1.070		
	рН	Scale 0-14	ASTM D1287*		7.69		
	Nitrites	ppm	Alcan Test Kit*		360		
	Reserve Alkalinity	Scale 0-20	ASTM D1121*		5.3		
	Percentage Glycol		ASTM D3321*		52.4		
	Freezing Point	°C	ASTM D3321*		-40		
	Carboxylate						
	Silicon	ppm	ASTM D5185(m)		56		
	Phosphorus	ppm	ASTM D5185(m)		29		
	Boron	ppm	ASTM D5185(m)		7		
	Molybdenum	ppm	ASTM D5185(m)		14		
	Sodium	ppm	ASTM D5185(m)		6812		
	Potassium Coolort Color	ppm	ASTM D5185(m)		448 Dink		
	Coolant Color		Visual*		Pink		





**CALA** ISO 17025:2017 Accredited Laboratory

Laboratory : WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9 Sample No. : LH0289510

Lab Number : 02637023 Unique Number : 5786185

Test Package : COOL ( Additional Tests: GlycolType )

To discuss this sample report, contact Customer Service at 1-800-268-2131. Test denoted (\*) outside scope of accreditation, (m) method modified, (e) tested at external lab. Validity of results and interpretation are based on the sample and information as supplied.

Received : 22 May 2024 **Tested** : 23 May 2024 : 23 May 2024 - Kevin Marson Diagnosed

> 12531992 ONTARIO INC. 64 ELLIOTT ROAD STERLING, ON CA K0K 3E0

> Contact: Service Manager

T:

F: