

WEAR	
CONTAMINATION	NORMAL
FLUID CONDITION	NORMAL

Locomotives

2003 Component Railway diesel Fluid RAILWAY ENGINE OIL SAE 40 (243 GAL)

RECOMMENDATION

Resample at the next service interval to monitor. Please contact your representative for information regarding the proper sampling kits for your service. NOTE: We recommend using MOB 3 test kits, this testkit includes Analytical Ferrography which provides a detailed morphological analysis of wear particles present in the fluid. this testkit includes BN to determine the suitability of the oil for continued use.

WEAR

Component wear rates appear to be normal (unconfirmed).

CONTAMINATION

There is no indication of any contamination in the oil.


.....


FLUID CONDITION


The condition of the oil is acceptable for the time in service (unconfirmed).

Test	UOM	Method	Limit/Abn	Current	History1	History2
Sample Number		Client Info		WC0938520	WC0938524	WC0938532
Sample Date		Client Info		11 Jun 2024	05 Jun 2024	29 May 2024
Machine Age	hrs	Client Info		0	0	0
Oil Age	hrs	Client Info		0	0	0
Filter Age	hrs	Client Info		0	0	0
Oil Changed		Client Info		Not Changd	Not Changd	Not Changd
Filter Changed		Client Info		Not Changd	Not Changd	Not Changd
Sample Status				NORMAL	NORMAL	NORMAL
Iron	ppm	ASTM D5185(m)	>100	17	16	16
Chromium	ppm	ASTM D5185(m)	>15	1	1	<1
Nickel	ppm	ASTM D5185(m)	>5	<1	<1	<1
Titanium	ppm	ASTM D5185(m)		0	0	0
Silver	ppm	ASTM D5185(m)	>2	<1	0	0
Aluminum	ppm	ASTM D5185(m)	>10	3	3	3
Lead	ppm	ASTM D5185(m)	>75	1	1	<1
Copper	ppm	ASTM D5185(m)	>90	8	8	8
Tin	ppm	ASTM D5185(m)	>30	<1	<1	<1
Vanadium	ppm	ASTM D5185(m)		0	0	0
Silicon	ppm	ASTM D5185(m)	>17	4	4	3
Potassium	ppm	ASTM D5185(m)	>20	2	<1	1
Fuel		WC Method	>4	<1.0	<1.0	<1.0
Water		WC Method	>0.20	NEG	NEG	NEG
Glycol		WC Method		NEG	NEG	NEG
Soot %	%	ASTM D7844*		0.2	0.2	0.2
Nitration	Abs/cm	ASTM D7624*	>20	12.1	12.1	12.2
Sulfation	Abs/.1mm	ASTM D7415*	>30	14.8	14.6	14.7
Emulsified Water	scalar	Visual*	>0.20	NEG	NEG	NEG
Sodium		ASTM D5185(m)		3	3	3
Boron	ppm	ASTM D5185(m)	10	1		
Barium	ppm ppm	ASTM D5185(m)	10	0	<1 0	<1 0
Molybdenum	ppm	ASTM D5185(m)	25	0	0	0
Manganese	ppm	ASTM D5185(m)	20	ہ <1	<1	<1
Magnesium		ASTM D5185(m)	20	18	17	17
Calcium	ppm ppm	ASTM D5185(m)	4500	4571	4721	4630
Phosphorus		ASTM D5185(m)	10	5	4	3
Zinc	ppm	ASTM D5185(m)	10	5	4	4
Sulfur	ppm	ASTM D5185(m)	5000	2934	4 2994	2856
Oxidation	ppm Abs/.1mm	ASTM D5165(III) ASTM D7414*	>25	2934 9.5	9.5	9.5
Visc @ 100°C	cSt	ASTM D7279(m)	14.4	14.9	14.8	14.8

Contact/Location: Richard Rochon - VALCOPTR







Laboratory : WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9 Vale - Transportation (Mobile Equipment) CALA Sample No. Received Transportation Department, (Services - Mobile Equipment) : WC0938520 : 21 Jun 2024 Lab Number : 02643323 COPPER CLIFF, ON Tested : 21 Jun 2024 ISO 17025:2017 Accredited Unique Number : 5800862 : 21 Jun 2024 - Wes Davis CA POM 1N0 Diagnosed Laboratory Test Package : MOB 1 Contact: Richard Rochon To discuss this sample report, contact Customer Service at 1-800-268-2131. richard.rochon@vale.com Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab. T: (705)682-6014 Validity of results and interpretation are based on the sample and information as supplied.

Contact/Location: Richard Rochon - VALCOPTR Page 2 of 2

F: