

TPC450LN-MP

Customer: PTRHTF40122

VAN DE RAA MEAT & MORE

BOLDERWEG 45

ALMERE, 1332 BA Netherlands

Attn: Wilbert Snijers

Tel:

E-Mail: w.snijers@klt.nl

System Information

System Volume: 1000 ltr

Bulk Operating Temp: 527F / 275C

Heating Source:

Blanket:

Fluid: PETRO CANADA PETRO-THERM

Make: WANSON


Sample Information

Lab No: 02265957 Analyst: Philip Riley Sample Date: 01/24/19 Received Date: 02/04/19 Completed: 02/05/19

Recommendation: GCD shows 2 off spec results related to the lower end of the temperature range of the fluid. If safely possible try and vent the system to remove lighter materials and see if this has an impact. Some activity on the distillation curve to support this. No other items for concern, fluid is otherwise good, so recommend a vent (under safe conditions/procedure) and re test at normal frequency

Comments: (GCD) 10% Distillation Point is abnormally low. (GCD) % < 335°C is marginally high.





Petro-Canada makes no representation or warranty of any kind, either express or implied, as to the accuracy or completeness of the analysis and assumes no responsibility and shall have no liability whatsoever with respect to such analysis, or a party's use of it. Petro-Canada is a division of HollyFrontier Corporation.