



IVAN MACIEL


Customer: PTRHTF60039	System Information	Sample Information
NOVONIX Anode Materials	System Volume: 30 Itr	Lab No: 02572806
353 Corporate Place	Bulk Operating Temp: 392F / 200C	Analyst: Jake Finn
Chattonooga, TN 37419 US	Heating Source:	Sample Date: 07/13/23
Attn: Ivan Maciel	Blanket:	Received Date: 07/27/23
Tel: (423)903-5884	Fluid: PETRO CANADA CALFLO AF	Completed: 07/31/23
E-Mail: ivan.maciel@novonixgroup.com	Make: USALAB	Jake Finn
		jake.finn@HFSinclair.com

Recommendation: Overall, sample is in great condition despite the dark color. Acid number was flagged just above the warning level, but Acid Number is best considered over time as a rate of change/trend instead of individual data points. GCD 90% distillation point was also flagged as marginally high but this value is not concerning when considering the other GCD values. Some very lite debris is noted by the lab which is not abnormal. No other concerning signs of degradation, wear or contamination. COC flash point is strong at 226°C. Test results indicate that the current fluid is suitable for continued use, but can be changed out when convenient within the next year if desired.

Comments: Acid Number (AN) is abnormally high. (GCD) 90% Distillation Point is marginally high.







Petro-Canada makes no representation or warranty of any kind, either express or implied, as to the accuracy or completeness of the analysis and assumes no responsibility and shall have no liability whatsoever with respect to such analysis, or a party's use of it. Petro-Canada is a division of HollyFrontier Corporation.

Report ID: [02572806] (Generated: 08/01/2023 12:27:21) - Page 2 - Copyright 2023 Wearcheck Inc. All Rights Reserved.