

OIL ANALYSIS REPORT

XAC012

Component Compressor

NOT GIVEN (--- GAL)

Sample Rating Trend



DIAGNOSIS

Recommendation

Resample at the next service interval to monitor. Please specify the brand, type, and viscosity of the oil on your next sample.

All component wear rates are normal.

Contamination

The water content is negligible. There is no indication of any contamination in the oil.

Fluid Condition

The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.

				Nov2022		
SAMPLE INFORM	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		PC0037972		
Sample Date		Client Info		21 Nov 2022		
Machine Age	hrs	Client Info		21000		
Oil Age	hrs	Client Info		900		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
WEAR METALS	S	method	limit/base	current	history1	history2
PQ		ASTM D8184*		0		
Iron	ppm	ASTM D5185(m)	>50	<1		
Chromium	ppm	ASTM D5185(m)	>10	0		
Nickel	ppm	ASTM D5185(m)		0		
Titanium	ppm	ASTM D5185(m)		0		
Silver	ppm	ASTM D5185(m)		0		
Aluminum	ppm	ASTM D5185(m)	>25	<1		
Lead	ppm	ASTM D5185(m)	>25	0		
Copper	ppm	ASTM D5185(m)	>50	<1		
Tin	ppm	ASTM D5185(m)	>15	0		
Antimony	ppm	ASTM D5185(m)		0		
Vanadium	ppm	ASTM D5185(m)		0		
Beryllium	ppm	ASTM D5185(m)		0		
Cadmium	ppm	ASTM D5185(m)		0		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185(m)		<1		
Barium	ppm	ASTM D5185(m)		0		
Molybdenum	ppm	ASTM D5185(m)		0		
Manganese	ppm	ASTM D5185(m)		0		
Magnesium	ppm	ASTM D5185(m)		<1		
Calcium	ppm	ASTM D5185(m)		0		
Phosphorus	ppm	ASTM D5185(m)		424		
Zinc	ppm	ASTM D5185(m)		4		
Sulfur	ppm	ASTM D5185(m)		31		
Lithium	ppm	ASTM D5185(m)		<1		
CONTAMINAN	TS	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185(m)	>25	<1		
Sodium	ppm	ASTM D5185(m)		<1		
Potassium	ppm	ASTM D5185(m)	>20	0		
Water	%	ASTM D6304*	>0.1	0.003		
ppm Water	ppm	ASTM D6304*	>1000	33.8		
FLUID DEGRAD	ATION	method	limit/base	current	history1	history2


Acid Number (AN)

mg KOH/g ASTM D974*

0.21



OIL ANALYSIS REPORT





ISO 17025:2017 Accredited

Laboratory

Sample No. Lab Number **Unique Number**

: PC0037972

Received : 02524200 Diagnosed

: Kevin Marson : 5489181 Diagnostician Test Package : IND 2 (Additional Tests: KF, KV100, VI)

: 23 Nov 2022

To discuss this sample report, contact Customer Service at 1-800-268-2131. Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab. Validity of results and interpretation are based on the sample and information as supplied.

415 ELMIRA RD. NORTH GUELPH, ON **CA N1K 1H3**

Contact: Dennis Kwok dennis.kwok@linamar.com T: (519)763-5369

F: (519)763-5302