

OIL ANALYSIS REPORT

Sample Rating Trend

WEAR

Area Infasco - 102400 A2308139

Component Unknown Component Fluid FORM AND LUBE (--- GAL)

DIAGNOSIS

A Recommendation

This is a baseline read-out on the submitted sample.

📥 Wear

Copper, nickel and iron ppm levels are noted.

Contamination

There is a high amount of silt present in the sample.


Fluid Condition

{not applicable}

SAMPLE INFORM	ATION	method	limit/base	current	history1	history2
Sample Number		Client Info		E30000050		
Sample Date		Client Info		23 Aug 2023		
Machine Age	hrs	Client Info		0		
Oil Age	hrs	Client Info		0		
Oil Changed		Client Info		N/A		
Sample Status				ATTENTION		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185(m)		162		
Chromium	ppm	ASTM D5185(m)		1		
Nickel	ppm	ASTM D5185(m)		11		
Titanium	ppm	ASTM D5185(m)		<1		
Silver	ppm	ASTM D5185(m)		<1		
Aluminum	ppm	ASTM D5185(m)		3		
Lead	ppm	ASTM D5185(m)		8		
Copper	ppm	ASTM D5185(m)		250		
Tin	ppm	ASTM D5185(m)		0		
Antimony		ASTM D5185(m)		0		
Vanadium	ppm	ASTM D5185(m) ASTM D5185(m)		ں <1		
	ppm	. ,				
Beryllium	ppm	ASTM D5185(m)		0		
Cadmium	ppm	ASTM D5185(m)		<1		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185(m)		2		
Barium	ppm	ASTM D5185(m)		0		
Molybdenum	ppm	ASTM D5185(m)		<1		
Manganese	ppm	ASTM D5185(m)		4		
Magnesium	ppm	ASTM D5185(m)		8		
Calcium	ppm	ASTM D5185(m)		394		
Phosphorus	ppm	ASTM D5185(m)		472		
Zinc	ppm	ASTM D5185(m)		307		
Sulfur	ppm	ASTM D5185(m)		3517		
Lithium	ppm	ASTM D5185(m)		<1		
CONTAMINANTS		method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185(m)		4		
Sodium	ppm	ASTM D5185(m)		4		
Potassium	ppm	ASTM D5185(m)	>20	2		
VISUAL		method	limit/base	current	history1	history2
White Metal	scalar	Visual*	NONE	VLITE		
Yellow Metal	scalar	Visual*	NONE	NONE		
Precipitate	scalar	Visual*	NONE	NONE		
Silt	scalar	Visual*	NONE			
	scalar	Visual*	NONE	LIGHT		
Debris						
Debris Sand/Dirt	scalar	Visual*	NONE	NONE		
	scalar scalar	Visual* Visual*	NONE NORML	NONE NORML		



OIL ANALYSIS REPORT



: WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9 Environmental 360 Solutions Ltd. Laboratory CALA Sample No. : E30000050 Received : 25 Aug 2023 640 Victoria Street Lab Number : 02578574 Diagnosed : 30 Aug 2023 Cobourg, ON ISO 17025:2017 Accredited Laboratory Unique Number : 5631634 Diagnostician : Tatiana Sorkina CA K9A 5H5 Test Package : TEST (Additional Tests: Bottom, ICP) Contact: Tatiana Sorkina To discuss this sample report, contact Customer Service at 1-800-268-2131. tsorkina@e360s.ca T: (800)263-3939 Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab. Validity of results and interpretation are based on the sample and information as supplied. F: (905)373-4950

Contact/Location: Tatiana Sorkina - CHECOB