




OIL ANALYSIS REPORT

Sample Rating Trend

NORMAL



Area
[510168]
 Machine Id
1919764
 Component
Unknown Component
 Fluid
NOT GIVEN (--- GAL)



DIAGNOSIS

Recommendation

This is a baseline read-out on the submitted sample.

Wear

{not applicable}

Contamination

ISO Cleanliness Code (ISO 4406:1999): 14/12/9;
 Cumulative particle counts >4µm = 112, >6µm = 33,
 >14µm = 3, >21µm = 1, >38µm = 0, >71µm = 0.

Fluid Condition

{not applicable}

SAMPLE INFORMATION

	method	limit/base	current	history1	history2
Sample Number	Client Info		PP	---	---
Sample Date	Client Info		25 Aug 2023	---	---
Machine Age	hrs	Client Info	0	---	---
Oil Age	hrs	Client Info	0	---	---
Oil Changed	Client Info		N/A	---	---
Sample Status			NORMAL	---	---



FLUID CLEANLINESS

	method	limit/base	current	history1	history2
Particles >4µm	ASTM D7647	>5000	112	---	---
Particles >6µm	ASTM D7647	>1300	33	---	---
Particles >14µm	ASTM D7647	>160	3	---	---
Particles >21µm	ASTM D7647	>40	1	---	---
Particles >38µm	ASTM D7647	>10	0	---	---
Particles >71µm	ASTM D7647	>3	0	---	---
Oil Cleanliness	ISO 4406 (c)	>19/17/14	14/12/9	---	---

VISUAL

	method	limit/base	current	history1	history2
White Metal	scalar	Visual*	NONE	NONE	---
Yellow Metal	scalar	Visual*	NONE	NONE	---
Precipitate	scalar	Visual*	NONE	NONE	---
Silt	scalar	Visual*	NONE	NONE	---
Debris	scalar	Visual*	NONE	NONE	---
Sand/Dirt	scalar	Visual*	NONE	NONE	---
Appearance	scalar	Visual*	NORML	NORML	---
Odor	scalar	Visual*	NORML	NORML	---




SAMPLE IMAGES

	method	limit/base	current	history1	history2
Color				no image	no image
Bottom				no image	no image



OIL ANALYSIS REPORT

GRAPHS



ISO 17025:2017
Accredited
Laboratory

Laboratory : WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9
Sample No. : PP **Received** : 28 Aug 2023
Lab Number : 02578940 **Diagnosed** : 29 Aug 2023
Unique Number : 5632000 **Diagnostician** : Kevin Marson
Test Package : TEST (Additional Tests: PrtCount)

To discuss this sample report, contact Customer Service at 1-800-268-2131.
 Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab.
 Validity of results and interpretation are based on the sample and information as supplied.

TESTMARK LABORATORIES
 6820 KITIMAT ROAD
 MISSISSAUGA, ON
 CA L5L 5T5
 Contact: Mississauga Sample
 mississauga.oils@testmark.ca
 T: (905)821-1112
 F: