


OIL ANALYSIS REPORT

Area [4376] 12-030-CD

Component **New (Unused) Oil** Fluid

PETRO CANADA LUMINOL LS B-4 (--- GAL)





Recommendation

This is a baseline read-out on the submitted sample.

Wear

{not applicable}

Contamination

{not applicable}

Fluid Condition

Pour Point = -60°C.

SAMPLE INFOR	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		PP0000869		
Sample Date		Client Info		22 Dec 2023		
Machine Age	hrs	Client Info		0		
Oil Age	hrs	Client Info		0		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
CONTAMINATIC	N	method	limit/base	current	history1	history2
Water		WC Method		NEG		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185(m)	>5	0		
Chromium	ppm	ASTM D5185(m)	>5	0		
Nickel	ppm	ASTM D5185(m)	>5	<1		
Titanium	ppm	ASTM D5185(m)		0		
Silver	ppm	ASTM D5185(m)	>5	0		
Aluminum	ppm	ASTM D5185(m)	>5	<1		
Lead	ppm	ASTM D5185(m)	>5	0		
Copper	ppm	ASTM D5185(m)	>5	0		
Tin	ppm	ASTM D5185(m)	>5	0		
Antimony	ppm	ASTM D5185(m)		0		
Vanadium	ppm	ASTM D5185(m)		0		
Beryllium	ppm	ASTM D5185(m)		0		
Cadmium	ppm	ASTM D5185(m)		0		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185(m)		0		
Barium	ppm	ASTM D5185(m)		0		
Molybdenum	ppm	ASTM D5185(m)		0		
Manganese	ppm	ASTM D5185(m)		0		
Magnesium	ppm	ASTM D5185(m)		0		
Calcium	ppm	ASTM D5185(m)		0		
Phosphorus	ppm	ASTM D5185(m)		0		
Zinc	ppm	ASTM D5185(m)		<1		
Sulfur	ppm	ASTM D5185(m)		0		
Lithium	ppm	ASTM D5185(m)		<1		
CONTAMINANT	S	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185(m)	>15	0		
Sodium	ppm	ASTM D5185(m)		0		
Potassium	ppm	ASTM D5185(m)	>20	1		
VISUAL		method	limit/base	current	history1	history2
White Metal	scalar	Visual*	NONE	NONE		
Yellow Metal	scalar	Visual*	NONE	NONE		
Precipitate	scalar	Visual*	NONE	NONE		
Silt	scalar	Visual*	NONE	NONE		
Debris	scalar	Visual*	NONE	NONE		
Sand/Dirt	scalar	Visual*	NONE	NONE		
Appearance	scalar	Visual*	NORML	NORML		
Odor	scalar	Visual*	NORML	NORML	ation:-Mark Mich	alkoff OLENIA



OIL ANALYSIS REPORT

FLUID PROPERTIES		method	limit/base	current	history1	history2
Pour Point	°C	ASTM D97*		-60		
SAMPLE IMAGES		method	limit/base	current	history1	history2
Color					no image	no image
Bottom					no image	no image
GRAPHS						



CALA ISO 17025:2017 Accredited Laboratory

: WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9 Laboratory Sample No. : PP0000869 Recieved : 27 Dec 2023 Lab Number : 02605486 Diagnosed : 27 Dec 2023 Unique Number : 5698571 Diagnostician : Kevin Marson Test Package : TEST (Additional Tests: ICP, PourPt, Visual)

To discuss this sample report, contact Customer Service at 1-800-268-2131. Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab. Validity of results and interpretation are based on the sample and information as supplied.

Oleo Energies 5800 Thorold Stone Road Niagara Falls, ON CA L2J 1A2 Contact: Mark Michalkoff

mmichalkoff@oleoenergies.com T: (905)358-5133

Contact/Location: Mark Michalkoff - OLENIA