

OIL ANALYSIS REPORT

Area [4399] 01-013-CE

Component **New (Unused) Oil** Fluid

PETRO CANADA LUMINOL TRI (--- GAL)

Sample Rating Trend NORMAL

Recommendation

This is a baseline read-out on the submitted sample.

Wear

{not applicable}

Contamination

{not applicable}

Fluid Condition

Pour Point = -60°C.

SAMPLE INFORM	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		PP0000882		
Sample Date		Client Info		12 Jan 2024		
Machine Age	hrs	Client Info		0		
Oil Age	hrs	Client Info		0		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
CONTAMINATIO	N	method	limit/base	current	history1	history2
Water		WC Method		NEG		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185(m)	>5	0		
Chromium	ppm	ASTM D5185(m)	>5	0		
Nickel	ppm	ASTM D5185(m)	>5	0		
Titanium	ppm	ASTM D5185(m)		0		
Silver	ppm	ASTM D5185(m)	>5	0		
Aluminum	ppm	ASTM D5185(m)	>5	<1		
Lead	ppm	ASTM D5185(m)	>5	0		
Copper	ppm	ASTM D5185(m)	>5	0		
Tin	ppm	ASTM D5185(m)	>5	0		
Antimony	ppm	ASTM D5185(m)		0		
Vanadium	ppm	ASTM D5185(m)		0		
Beryllium	ppm	ASTM D5185(m)		0		
Cadmium	ppm	ASTM D5185(m)		0		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185(m)	0	0		
Barium	ppm	ASTM D5185(m)	0	0		
Molybdenum	ppm	ASTM D5185(m)	0	0		
Manganese	ppm	ASTM D5185(m)	0	0		
Magnesium	ppm	ASTM D5185(m)	0	<1		
Calcium	ppm	ASTM D5185(m)		0		
Phosphorus	ppm	ASTM D5185(m)	0	0		
Zinc	ppm	ASTM D5185(m)		<1		
Sulfur	ppm	ASTM D5185(m)	10	0		
Lithium	ppm	ASTM D5185(m)		<1		
CONTAMINANTS	;	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185(m)	>15	0		
Sodium	ppm	ASTM D5185(m)	>5	0		
Potassium			- 0	U		
	ppm	ASTM D5185(m)	>20	<1		
VISUAL	ppm	ASTM D5185(m) method			history1	history2
VISUAL White Metal	ppm		>20	<1		
White Metal	scalar	method Visual*	>20 limit/base NONE	<1 current NONE	history1	history2
White Metal Yellow Metal	scalar scalar	method Visual* Visual*	>20 limit/base NONE NONE	<1 current NONE NONE	history1	history2
White Metal Yellow Metal Precipitate	scalar scalar scalar	method Visual* Visual* Visual*	>20 limit/base NONE NONE NONE	current NONE NONE NONE	history1 	history2
White Metal Yellow Metal Precipitate Silt	scalar scalar scalar scalar	method Visual* Visual* Visual* Visual*	>20 limit/base NONE NONE NONE NONE	current NONE NONE NONE NONE NONE	history1	history2
White Metal Yellow Metal Precipitate Silt Debris	scalar scalar scalar scalar scalar	method Visual* Visual* Visual* Visual* Visual*	>20 limit/base NONE NONE NONE NONE NONE NONE	current NONE NONE NONE NONE NONE NONE	history1	history2
White Metal Yellow Metal Precipitate Silt Debris Sand/Dirt	scalar scalar scalar scalar scalar scalar	method Visual* Visual* Visual* Visual* Visual* Visual*	>20 limit/base NONE NONE NONE NONE NONE NONE NONE	current NONE NONE NONE NONE NONE NONE NONE NONE	history1	history2
White Metal Yellow Metal Precipitate Silt Debris	scalar scalar scalar scalar scalar	method Visual* Visual* Visual* Visual* Visual*	>20 limit/base NONE NONE NONE NONE NONE NONE	current NONE NONE NONE NONE NONE NONE	history1	history2



OIL ANALYSIS REPORT

FLUID PROPERTIES		method	limit/base	current	history1	history2
Pour Point	°C	ASTM D97*	-60	-60		
SAMPLE IMAGES		method	limit/base	current	history1	history2
Color					no image	no image
Bottom					no image	no image
GRAPHS						



CALA ISO 17025:2017 Accredited Laboratory

: WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9 Laboratory Sample No. : PP0000882 Recieved : 15 Jan 2024 Lab Number : 02608825 Diagnosed : 15 Jan 2024 Unique Number : 5709911 Diagnostician : Kevin Marson Test Package : TEST (Additional Tests: ICP, PourPt, Visual)

To discuss this sample report, contact Customer Service at 1-800-268-2131. Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab. Validity of results and interpretation are based on the sample and information as supplied.

Oleo Energies 5800 Thorold Stone Road Niagara Falls, ON CA L2J 1A2 Contact: Mark Michalkoff

mmichalkoff@oleoenergies.com T: (905)358-5133

Contact/Location: Mark Michalkoff - OLENIA