



OIL ANALYSIS REPORT

Sample Rating Trend

NORMAL



Machine Id
SYN BLEND HYD 46
Component
Unknown Component
Fluid
SYN BLEND HYD 46 (--- GAL)



DIAGNOSIS

Recommendation

This is a baseline read-out on the submitted sample.

Fluid Condition

Pour Point = -42°C.

SAMPLE INFORMATION

	method	limit/base	current	history1	history2
Sample Number	Client Info		PP	---	---
Sample Date	Client Info		11 Mar 2024	---	---
Machine Age	hrs	Client Info	0	---	---
Oil Age	hrs	Client Info	0	---	---
Oil Changed	Client Info		N/A	---	---
Sample Status			NORMAL	---	---

CONTAMINATION

	method	limit/base	current	history1	history2
Water	WC Method		NEG	---	---



VISUAL

	method	limit/base	current	history1	history2
White Metal	scalar	Visual*	NONE	---	---
Yellow Metal	scalar	Visual*	NONE	---	---
Precipitate	scalar	Visual*	NONE	---	---
Silt	scalar	Visual*	NONE	---	---
Debris	scalar	Visual*	NONE	---	---
Sand/Dirt	scalar	Visual*	NONE	---	---
Appearance	scalar	Visual*	NORML	---	---
Odor	scalar	Visual*	NORML	---	---

FLUID PROPERTIES

	method	limit/base	current	history1	history2
Pour Point	°C	ASTM D97*	-42	---	---

SAMPLE IMAGES

	method	limit/base	current	history1	history2
Color				no image	no image
Bottom				no image	no image



OIL ANALYSIS REPORT

GRAPHS



ISO 17025:2017
Accredited
Laboratory

Laboratory : WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9
Sample No. : PP
Lab Number : 02621640
Unique Number : 5746759
Test Package : TEST (Additional Tests: PourPt)

Received : 12 Mar 2024
Tested : 12 Mar 2024
Diagnosed : 13 Mar 2024 - Kevin Marson

IRVING BLENDING & PACKAGING
P.O.BOX 1169, 555 COURTENAY CAUSEWAY
Saint John, NB
CA E2L 4E6

Contact: Blaine Sherwood
Blaine.Sherwood@irvingoil.com

T: (506)202-7730
F: (506)632-7001

To discuss this sample report, contact Customer Service at 1-800-268-2131.

Test denoted (*) outside scope of accreditation, (m) method modified, (e) tested at external lab.

Validity of results and interpretation are based on the sample and information as supplied.