

## **OIL ANALYSIS REPORT**

Sample Rating Trend



Machine Id FLA-PSFD PD-032824-6

Component Unknown Component {not provided} (--- GAL)

## Recommendation


This is a baseline read-out on the submitted sample.

				Mar2024		
SAMPLE INFORI	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		PP		
Sample Date		Client Info		31 Mar 2024		
Machine Age	hrs	Client Info		0		
Oil Age	hrs	Client Info		0		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
CONTAMINATIO	N	method	limit/base	current	history1	history2
Water		WC Method		NEG		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185(m)		0		
Chromium	ppm	ASTM D5185(m)		0		
Nickel	ppm	ASTM D5185(m)		0		
Titanium	ppm	ASTM D5185(m)		0		
Silver	ppm	ASTM D5185(m)		0		
Aluminum	ppm	ASTM D5185(m)		0		
Lead	ppm	ASTM D5185(m)		0		
Copper	ppm	ASTM D5185(m)		0		
Tin	ppm	ASTM D5185(m)		0		
Antimony	ppm	ASTM D5185(m)		0		
Vanadium	ppm	ASTM D5185(m)		0		
Beryllium	ppm	ASTM D5185(m)		0		
Cadmium	ppm	ASTM D5185(m)		0		
	ррпп		1114-//			
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185(m)		<1		
Barium	ppm	ASTM D5185(m)		0		
Molybdenum	ppm	ASTM D5185(m)		0		
Manganese	ppm	ASTM D5185(m)		0		
Magnesium	ppm	ASTM D5185(m)		2		
Calcium	ppm	ASTM D5185(m)		55		
Phosphorus	ppm	ASTM D5185(m)		314		
Zinc	ppm	ASTM D5185(m)		413		
Sulfur	ppm	ASTM D5185(m)		718		
Lithium	ppm	ASTM D5185(m)		<1		
CONTAMINANTS	3	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185(m)		0		
Sodium	ppm	ASTM D5185(m)		0		
Potassium	ppm	ASTM D5185(m)	>20	<1		
VISUAL		method	limit/base	current	history1	history2
White Metal	scalar	Visual*	NONE	NONE		
Yellow Metal	scalar	Visual*	NONE	NONE		
Precipitate	scalar	Visual*	NONE	NONE		
Silt	scalar	Visual*	NONE	NONE		
Debris	scalar	Visual*	NONE	NONE		
Sand/Dirt	scalar	Visual*	NONE	VLITE		
Appearance	scalar	Visual*	NORML	NORML		
Odor	scalar	Visual*	NORML		IDI <u>LE</u> INGARTI	NER_FORHA
J J J J	oodial	riodai	. TOT HVIL			Dogo 1 of



## **OIL ANALYSIS REPORT**

**Bottom** 



FLUID PROPERT	IES	method	limit/base	current	history1	history2
Visc @ 40°C	cSt	ASTM D7279(m)		43.4		
Visc @ 100°C	cSt	ASTM D7279(m)		7.0		
Viscosity Index (VI)	Scale	ASTM D2270*		119		
Pour Point	°C	ASTM D97*		-36		
COC Flash Point	°C	ASTM D92*		255		
SAMPLE IMAGES	\$	method	limit/base	current	history1	history2
SAMPLE IMAGES		method	limit/base	current	no image	history2 no image



CALA ISO 17025:2017 Accredited Laboratory

Laboratory Sample No.

: PP **Lab Number** : 02625872 Unique Number : 5759004

: WearCheck - C8-1175 Appleby Line, Burlington, ON L7L 5H9

Received **Tested** Diagnosed

: 01 Apr 2024 : 08 Apr 2024

120 CHATHAM ST. HAMILTON, ON CA L8P 2B5

T: (905)525-7192

F: (905)525-7024

FORSYTHE LUBRICATION

no image

: 08 Apr 2024 - Kevin Marson Test Package : TEST ( Additional Tests: COC Flash, ICP, KV100, KV40, PourPt, VI ) Contact: HEIDI LEINGARTNER

To discuss this sample report, contact Customer Service at 1-800-268-2131. Test denoted (\*) outside scope of accreditation, (m) method modified, (e) tested at external lab.

Validity of results and interpretation are based on the sample and information as supplied.