




COOLANT REPORT

Sample Rating Trend

NORMAL



Machine Id
1621001 - PINNACLE DTM ISO 680

Component
New (Unused) Oil
Fluid
{not provided} (--- GAL)

DIAGNOSIS

Recommendation

All tests and evaluation performed at subcontracted ISO 17025 laboratory. ASTM D130 Copper Strip Corrosion: 1A for 3 hours at 212F (100C). ASTM D1401 Water Separability: 42/37/2. Please note that this is an updated copy for ASTM D93 PMCC measured at 266.5C.

SAMPLE INFORMATION

	method	limit/base	current	history1	history2
Sample Number	Client Info		WC05723603	---	---
Sample Date	Client Info		20 Dec 2022	---	---
Machine Age	hrs	Client Info	0	---	---
Oil Age	hrs	Client Info	0	---	---
Oil Changed	Client Info		N/A	---	---
Sample Status			NORMAL	---	---




VISUAL

	method	limit/base	current	history1	history2
Color			<i>no image</i>	<i>no image</i>	<i>no image</i>
Bottom			<i>no image</i>	<i>no image</i>	<i>no image</i>



COOLANT REPORT

GRAPHS



Certificate L2367

Laboratory : WearCheck USA - 501 Madison Ave., Cary, NC 27513
Sample No. : WC05723603 **Received** : 21 Dec 2022
Lab Number : 05723603 **Diagnosed** : 09 Jan 2023
Unique Number : 10268184 **Diagnostician** : Doug Bogart
Test Package : TEST (Additional Tests: CC FLASH, COOL, KF)

PETROLEUM TECHNOLOGIES GROUP
 4665 BROADMOOR AVE, SUITE 150
 GRAND RAPIDS, MI
 US 49512
 Contact: JAMES KRAFT
 james@oil-lab.com
 T: (616)698-9399
 F:

To discuss this sample report, contact Customer Service at 1-800-237-1369.

* - Denotes test methods that are outside of the ISO 17025 scope of accreditation.

Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)