

OIL ANALYSIS REPORT

ROTO XTEND ATLAS COPCO AOF235346 - NTN BOWERS (S/N APF235346)

Compressor



Sample Rating Trend



Recommendation

Resample at the next service interval to monitor.

All component wear rates are normal.

Contamination

There is no indication of any contamination in the


Fluid Condition

The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.

				Mar2023		
SAMPLE INFORM	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		UCH05818944		
Sample Date		Client Info		29 Mar 2023		
Machine Age	hrs	Client Info		28808		
Oil Age	hrs	Client Info		8000		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
CONTAMINATION		method	limit/base	current	history1	history2
Water		WC Method	>0.1	NEG		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185m	>50	0		
Chromium	ppm	ASTM D5185m	>5	0		
Nickel	ppm	ASTM D5185m		0		
Titanium	ppm	ASTM D5185m		0		
Silver	ppm	ASTM D5185m		0		
Aluminum	ppm	ASTM D5185m	>15	0		
Lead	ppm	ASTM D5185m	>65	0		
Copper	ppm	ASTM D5185m	>65	<1		
Tin	ppm	ASTM D5185m	>10	0		
Vanadium	ppm	ASTM D5185m		0		
Cadmium	ppm	ASTM D5185m		0		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185m		0		
Barium	ppm	ASTM D5185m		0		
Molybdenum	ppm	ASTM D5185m		0		
Manganese	ppm	ASTM D5185m		<1		
Magnesium	ppm	ASTM D5185m		<1		
Calcium	ppm	ASTM D5185m		0		
Phosphorus	ppm	ASTM D5185m		11		
Zinc	ppm	ASTM D5185m		0		
Sulfur	ppm	ASTM D5185m		28		
CONTAMINANTS	;	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185m	>35	2		
Sodium	ppm	ASTM D5185m		<1		
Potassium	ppm	ASTM D5185m	>20	0		
FLUID DEGRADA	ATION	method	limit/base	current	history1	history2
Acid Number (AN)	mg KOH/g	ASTM D8045	0.14	0.617		



OIL ANALYSIS REPORT







Certificate L2367

Laboratory Sample No. Lab Number Unique Number

: 05818944 : 10427027 Test Package : IND 2

To discuss this sample report, contact Customer Service at 1-800-237-1369.

. ₹3 45

: WearCheck USA - 501 Madison Ave., Cary, NC 27513 : 13 Apr 2023 : UCH05818944 Recieved Diagnosed : 16 Apr 2023

Diagnostician

: Don Baldridge

Mar29/23

(mg K0H/

HTE TECHNOLOGIES 2021 CONGRESSIONAL DR ST LOUIS, MO US 63146

Contact: JOSH DYRLAND josh.dyrland@htetech.com T: (314)731-4444

* - Denotes test methods that are outside of the ISO 17025 scope of accreditation. Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)

F: (314)731-4424