


OIL ANALYSIS REPORT

Sample Rating Trend



NORMAL



CM-1 (S/N P12CTM1) Component

Gearbox

ROYAL PURPLE SYNFILM GT 220 (--- QTS)

Recommendation

Resample at the next service interval to monitor.

All component wear rates are normal.

Contamination

There is no indication of any contamination in the

Fluid Condition

The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.

,			Dec2020	Jun2023		
SAMPLE INFORM	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		RP0030075	RP0013484	
Sample Date		Client Info		05 Jun 2023	15 Dec 2020	
Machine Age	mths	Client Info		0	0	
Oil Age	mths	Client Info		6	4380	
Oil Changed		Client Info		Not Changd	Not Changd	
Sample Status				NORMAL	NORMAL	
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185m	>200	8	4	
Chromium	ppm	ASTM D5185m	>15	0	0	
Nickel	ppm	ASTM D5185m	>15	<1	<1	
Titanium	ppm	ASTM D5185m		0	0	
Silver	ppm	ASTM D5185m		0	0	
Aluminum	ppm	ASTM D5185m	>25	0	0	
Lead	ppm	ASTM D5185m	>100	0	0	
Copper	ppm	ASTM D5185m	>200	<1	<1	
Tin	ppm	ASTM D5185m	>25	0	0	
Antimony	ppm	ASTM D5185m	>5		0	
Vanadium	ppm	ASTM D5185m		0	0	
Cadmium	ppm	ASTM D5185m		0	0	
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185m		0	<1	
Barium	ppm	ASTM D5185m		2	0	
Molybdenum	ppm	ASTM D5185m		0	0	
Manganese	ppm	ASTM D5185m		<1	0	
Magnesium	ppm	ASTM D5185m		97	81	
Calcium	ppm	ASTM D5185m		4	3	
Phosphorus	ppm	ASTM D5185m		2	3	
Zinc	ppm	ASTM D5185m		9	0	
CONTAMINANTS	3	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185m	>50	3	3	
Sodium	ppm	ASTM D5185m		<1	1	
Potassium	ppm	ASTM D5185m	>20	<1	0	
Water	%	ASTM D6304	>0.2	0.020	0.014	
ppm Water	ppm	ASTM D6304	>2000	209.1	140.2	
FLUID DEGRADA	NOITA	method	limit/base	current	history1	history2

0.42


Acid Number (AN)

mg KOH/g ASTM D8045

0.399



OIL ANALYSIS REPORT







Laboratory Sample No. Lab Number

Unique Number

: 10519021 Test Package : IND 2

: WearCheck USA - 501 Madison Ave., Cary, NC 27513 : 14 Jun 2023 : RP0030075 Received : 05873918 Diagnosed : 16 Jun 2023 Diagnostician : Don Baldridge

To discuss this sample report, contact Customer Service at 1-800-237-1369.

* - Denotes test methods that are outside of the ISO 17025 scope of accreditation. Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012) **VEOLIA SMITHFIELD** P.O. BOX 17249 SMITHFIELD, RI

US 02917 Contact: KEVIN REGAN kevin.regan@veolia.com

T: (401)231-1506 F: (401)231-7089