

OIL ANALYSIS REPORT

Sample Rating Trend

NORMAL

BITLER PH2 22SC2 (S/N 07L400054) Component

Screw Compressor SOLEST 170 (3 GAL)

Recommendation

Resample at the next service interval to monitor.

Wear

All component wear rates are normal.

Contamination

There is no indication of any contamination in the oil.

Fluid Condition

The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.

SAMPLE INFORM	IATION	method	limit/base	current	history1	history2
Sample Number		Client Info		WC0812034	WC0689510	WC0613581
Sample Date		Client Info		25 Aug 2023	15 Aug 2022	10 Sep 2021
Machine Age	hrs	Client Info		0	0	0
Oil Age	hrs	Client Info		0	0	0
Oil Changed		Client Info		N/A	N/A	N/A
Sample Status				NORMAL	NORMAL	NORMAL
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185m	>60	<1	<1	0
Chromium	ppm	ASTM D5185m	>4	0	0	0
Nickel	ppm	ASTM D5185m		0	0	0
Titanium	ppm	ASTM D5185m		0	0	0
Silver	ppm	ASTM D5185m		0	0	<1
Aluminum	ppm	ASTM D5185m	>5	0	<1	0
Lead	ppm	ASTM D5185m	>10	0	<1	0
Copper	ppm	ASTM D5185m	>30	0	<1	0
Tin	ppm	ASTM D5185m	>15	0	1	<1
Antimony	ppm	ASTM D5185m				0
Vanadium	ppm	ASTM D5185m		0	0	0
Cadmium	ppm	ASTM D5185m		0	0	0
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185m		0	1	<1
Barium	ppm	ASTM D5185m		2	0	0
Molybdenum	ppm	ASTM D5185m		0	0	0
Manganese	ppm	ASTM D5185m		0	0	0
Magnesium	ppm	ASTM D5185m		0	0	0
Calcium	ppm	ASTM D5185m		0	0	0
Phosphorus	ppm	ASTM D5185m		3	5	3
Zinc	ppm	ASTM D5185m		0	0	0
Sulfur	ppm	ASTM D5185m		0	0	0
CONTAMINANTS		method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185m	>50	0	<1	0
Sodium	ppm	ASTM D5185m		0	<1	0
Potassium	ppm	ASTM D5185m	>20	<1	0	0
Water	%	ASTM D6304	>0.1	0.015	0.009	0.016
ppm Water	ppm	ASTM D6304	>1000	153.0	90.7	161.6
FLUID DEGRADA	TION	method	limit/base	current	history1	history2
Acid Number (AN)	mg KOH/g	ASTM D8045		0.015	0.013	0.015



OIL ANALYSIS REPORT



COMMERCIAL AIR SYSTEMS : WearCheck USA - 501 Madison Ave., Cary, NC 27513 10 COUNTY LINE RD BRANCHBURG, NJ US 08876 Contact: LARRY DELGUERCIO larry@commairsystems.com T: (908)927-9090 F: (908)921-9091

Aug15/22 -



Diagnostician : Doug Bogart Unique Number : 10629626 Test Package : IND 2 (Additional Tests: KF) Certificate L2367 To discuss this sample report, contact Customer Service at 1-800-237-1369. * - Denotes test methods that are outside of the ISO 17025 scope of accreditation. Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)

Received

Diagnosed

: 30 Aug 2023

: 31 Aug 2023

: WC0812034

: 05939014

Laboratory

Sample No.

Lab Number

Contact/Location: LARRY DELGUERCIO - COMBRANJ

NONE

NONE

NONE

NONE

NONE

NONE

NORML

NORML

NEG

NEG

148

NONE

NONE

NONE

NONE

NONE

NONE

NORML

NORML

NEG

NEG

155