

OIL ANALYSIS REPORT

Sample Rating Trend

NORMAL

ELECTRIC MOTOR EAST COOLING TOWER MOTOR

Top Bearing

ROYAL PURPLE SYNFILM 68 (5 GAL)

DIAGNOSIS

Recommendation

Resample at the next service interval to monitor.

Wear

All component wear rates are normal.

Contamination

The water content is negligible. There is no indication of any contamination in the oil.


Fluid Condition


The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.

SAMPLE INFORM	IATION	method	limit/base	current	history1	history2
Sample Number		Client Info		RP0035907	RP154736	
Sample Date		Client Info		04 Jan 2024	27 Sep 2018	
Machine Age	yrs	Client Info		0	0	
Oil Age	yrs	Client Info		0	2	
Oil Changed		Client Info		N/A	Changed	
Sample Status				NORMAL	NORMAL	
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185m	>20	2	<1	
Chromium	ppm	ASTM D5185m	>20	<1	0	
Nickel	ppm	ASTM D5185m	>20	0	<1	
Titanium	ppm	ASTM D5185m		<1	0	
Silver	ppm	ASTM D5185m		0	0	
Aluminum	ppm	ASTM D5185m	>20	2	0	
Lead	ppm	ASTM D5185m	>20	<1	0	
Copper	ppm	ASTM D5185m	>20	<1	<1	
Tin	ppm	ASTM D5185m	>20	<1	0	
Antimony	ppm	ASTM D5185m			0	
Vanadium	ppm	ASTM D5185m		0	0	
Cadmium	ppm	ASTM D5185m		0	0	
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185m		0	<1	
Barium	ppm	ASTM D5185m		0	0	
Molybdenum	ppm	ASTM D5185m		0	0	
Manganese	ppm	ASTM D5185m		0	<1	
Magnesium	ppm	ASTM D5185m	90	0	84	
Calcium	ppm	ASTM D5185m		112	2	
Phosphorus	ppm	ASTM D5185m		916	1	
Zinc	ppm	ASTM D5185m		202	<1	
CONTAMINANTS	1	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185m	>15	<1	1	
Sodium	ppm	ASTM D5185m		<1	<1	
Potassium	ppm	ASTM D5185m	>20	1	2	
Water	%	ASTM D6304	>2	0.016	0.037	
ppm Water	ppm	ASTM D6304		167	370	
FLUID DEGRADA	TION	method	limit/base	current	history1	history2
Acid Number (AN)	mg KOH/g	ASTM D8045	0.25	0.42	0.414	



OIL ANALYSIS REPORT






Viscosity @ 40°C

80 Abnorma

7 cSt (40°C)

> 65 Ab 60 Sep27/

Ba



: 16 Jan 2024

: 18 Jan 2024



Unique Number : Jonathan Hester : 10832647 Diagnostician Test Package : IND 2 Contact: CHARLES LOVE Certificate L2367 To discuss this sample report, contact Customer Service at 1-800-237-1369. charles_love@oxy.com * - Denotes test methods that are outside of the ISO 17025 scope of accreditation. Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)

Recieved

Diagnosed

Laboratory

Sample No.

Lab Number

: RP0035907

:06061265

Contact/Location: CHARLES LOVE - OXYDEE

1000 TIDAL ROAD

DEER PARK, TX

T: (281)476-2277

F: (281)476-2604

US 77536