


OIL ANALYSIS REPORT

SYNOIL 8K FG Machine Id QUINCY QGV-100 UN115437 - MSI EXPRESS

Component

Compressor





DIAGNOSIS

Recommendation

Resample at the next service interval to monitor.

Wear

All component wear rates are normal.

Contamination

There is no indication of any contamination in the oil


Fluid Condition

The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.

SAMPLE INFORM	MATION	method	limit/base	current	history1	history2
Sample Number		Client Info		UCZ06074900		
Sample Date		Client Info		20 Jan 2024		
Machine Age	hrs	Client Info		110674		
Oil Age	hrs	Client Info		2851		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
CONTAMINATIO	N	method	limit/base	current	history1	history2
Water		WC Method	>0.1	NEG		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185m	>50	0		
Chromium	ppm	ASTM D5185m	>5	0		
Nickel	ppm	ASTM D5185m		<1		
Titanium	ppm	ASTM D5185m		0		
Silver	ppm	ASTM D5185m		0		
Aluminum	ppm	ASTM D5185m	>15	1		
Lead	ppm	ASTM D5185m	>65	0		
Copper	ppm	ASTM D5185m	>65	0		
Tin	ppm	ASTM D5185m	>10	0		
Vanadium	ppm	ASTM D5185m		0		
Cadmium	ppm	ASTM D5185m		0		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185m	1	0		
Barium	ppm	ASTM D5185m	0.3	0		
Molybdenum	ppm	ASTM D5185m	0	0		
Manganese	ppm	ASTM D5185m	0	<1		
Magnesium	ppm	ASTM D5185m	0	<1		
Calcium	ppm	ASTM D5185m	0.5	<1		
Phosphorus	ppm	ASTM D5185m	536	103		
Zinc	ppm	ASTM D5185m	0.2	0		
Sulfur	ppm	ASTM D5185m	649	323		
CONTAMINANTS	3	method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185m	>35	4		
Sodium	ppm	ASTM D5185m		<1		
Potassium	ppm	ASTM D5185m	>20	0		
FLUID DEGRADA	ATION	method	limit/base	current	history1	history2
Acid Number (AN)	mg KOH/g	ASTM D8045	0.337	0.38		



OIL ANALYSIS REPORT







Certificate L2367

Laboratory Sample No. Lab Number Unique Number

: UCZ06074900 : 06074900

₹ 40

35

: 10856991 Test Package : IND 2

: WearCheck USA - 501 Madison Ave., Cary, NC 27513 : 30 Jan 2024 Recieved

Diagnosed : 01 Feb 2024 : Don Baldridge Diagnostician

To discuss this sample report, contact Customer Service at 1-800-237-1369.

Viscosity @ 40°C

* - Denotes test methods that are outside of the ISO 17025 scope of accreditation.

ZORN COMP & EQUIPMENT CO (GB)

Acid Number

(mg KOH/

733 POTTS AVE GREEN BAY, WI US 54304

Contact: DEAN SCHAD dean.schad@zornair.com

T: (920)391-8121 F: (920)499-1168

Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)