


OIL ANALYSIS REPORT



NORMAL



TOTE 21

Component New (Unused) Oil

{not provided} (--- GAL)

Recommendation

This is a baseline read-out on the submitted sample.


Fluid Condition

Viscosity of sample indicates oil is within 10W30 range. Non-engine oil, possible a UTF.

| | | | | Feb 2024 | | |
|---------------------------------|-----------------|---------------------------|-------------------|--------------|----------|----------|
| SAMPLE INFORM | MATION | method | limit/base | current | history1 | history2 |
| Sample Number | | Client Info | | TLC0000421 | | |
| Sample Date | | Client Info | | 16 Feb 2024 | | |
| Machine Age | hrs | Client Info | | 0 | | |
| Oil Age | hrs | Client Info | | 0 | | |
| Oil Changed | | Client Info | | N/A | | |
| Sample Status | | | | NORMAL | | |
| WEAR METALS | | method | limit/base | current | history1 | history2 |
| Iron | ppm | ASTM D5185m | >5 | 0 | | |
| Chromium | ppm | ASTM D5185m | >5 | 0 | | |
| Nickel | ppm | ASTM D5185m | >5 | 0 | | |
| Titanium | ppm | ASTM D5185m | | <1 | | |
| Silver | ppm | ASTM D5185m | >5 | 0 | | |
| Aluminum | ppm | ASTM D5185m | >5 | 0 | | |
| Lead | ppm | ASTM D5185m | >5 | <1 | | |
| Copper | ppm | ASTM D5185m | >5 | 0 | | |
| Tin | ppm | ASTM D5185m | >5 | 0 | | |
| Vanadium | ppm | ASTM D5185m | | 0 | | |
| Cadmium | ppm | ASTM D5185m | | 0 | | |
| ADDITIVES | | method | limit/base | current | history1 | history2 |
| Boron | ppm | ASTM D5185m | | 33 | | |
| Barium | ppm | ASTM D5185m | | 0 | | |
| Molybdenum | ppm | ASTM D5185m | | 0 | | |
| Manganese | ppm | ASTM D5185m | | 0 | | |
| Magnesium | ppm | ASTM D5185m | | 2 | | |
| Calcium | ppm | ASTM D5185m | | 393 | | |
| Phosphorus | ppm | ASTM D5185m | | 288 | | |
| Zinc | ppm | ASTM D5185m | | 148 | | |
| Sulfur | ppm | ASTM D5185m | | 6221 | | |
| CONTAMINANTS | ; | method | limit/base | current | history1 | history2 |
| Silicon | ppm | ASTM D5185m | >15 | 1 | | |
| Sodium | ppm | ASTM D5185m | | <1 | | |
| Codiaiii | ρρ | | | | | |
| | ppm | ASTM D5185m | >20 | <1 | | |
| Potassium | | ASTM D5185m ASTM D6304 | >20 | <1 0.023 | | |
| Potassium Water | ppm | | >20 | | | |
| Potassium Water | ppm % ppm | ASTM D6304 | >20 limit/base | 0.023 | | |
| Potassium Water ppm Water | ppm % ppm | ASTM D6304 ASTM D6304 | | 0.023 239 | | |



OIL ANALYSIS REPORT







Laboratory Sample No.

: TLC0000421 Lab Number : 06099510

Unique Number: 10897740

Received **Tested**

Diagnosed Test Package: FLEET (Additional Tests: FT-IR, ICP-NewOil, KF, KV100, PrtCount, VI)Contact: MICHAEL JACKSON

: 28 Feb 2024

: 23 Feb 2024

: 28 Feb 2024 - Doug Bogart

115 EMPIRE WAY ATLANTA, GA US 30354

To discuss this sample report, contact Customer Service at 1-800-237-1369.

mjackson@supplypro1.com T: (470)991-1693

* - Denotes test methods that are outside of the ISO 17025 scope of accreditation. Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)