

Area

## **OIL ANALYSIS REPORT**

Lubrication Room Purity FG Synthetic 220 Drum of Oil

Fluid GEAR OIL (PAO) ISO 220 (55 GAL)

#### Recommendation


This is a baseline read-out on the submitted sample. ( Customer Sample Comment: New oil )

### Contamination

There is no indication of any contamination in the oil. The amount and size of particulates present in the system are acceptable.

### Fluid Condition

The AN level is acceptable for this fluid. The condition of the oil is suitable for further service.





NORMAL

SAMPLE INFORM	IATION	method	limit/base	current	history1	history2
Sample Number		Client Info		WC0859498		
Sample Date		Client Info		28 May 2024		
Machine Age	hrs	Client Info		0		
Oil Age	hrs	Client Info		0		
Oil Changed		Client Info		N/A		
Sample Status				NORMAL		
CONTAMINATION	N	method	limit/base	current	history1	history2
Water		WC Method		NEG		
WEAR METALS		method	limit/base	current	history1	history2
Iron	ppm	ASTM D5185m		0		
Chromium	ppm	ASTM D5185m		0		
Nickel	ppm	ASTM D5185m		0		
Titanium	ppm	ASTM D5185m		0		
Silver	ppm	ASTM D5185m		0		
Aluminum	ppm	ASTM D5185m		0		
Lead	ppm	ASTM D5185m		0		
Copper	ppm	ASTM D5185m		0		
Tin	ppm	ASTM D5185m		0		
Vanadium	ppm	ASTM D5185m		0		
Cadmium	ppm	ASTM D5185m		0		
ADDITIVES		method	limit/base	current	history1	history2
Boron	ppm	ASTM D5185m	25	0		
Barium	ppm	ASTM D5185m	12	0		
Molybdenum	ppm	ASTM D5185m	5	0		
Manganese	ppm	ASTM D5185m		0		
Magnesium	ppm	ASTM D5185m	25	0		
Calcium	ppm	ASTM D5185m	25	0		
Phosphorus	ppm	ASTM D5185m	375	512		
Zinc	ppm	ASTM D5185m	25	0		
Sulfur	ppm	ASTM D5185m	4900	1271		
CONTAMINANTS		method	limit/base	current	history1	history2
Silicon	ppm	ASTM D5185m		<1		
Sodium	ppm	ASTM D5185m		<1		
Potassium	ppm	ASTM D5185m	>20	0		
FLUID CLEANLIN	IESS	method	limit/base	current	history1	history2
Particles >4µm		ASTM D7647	>5000	1219		
Particles >6µm		ASTM D7647	>1300	435		
Particles >14µm		ASTM D7647	>160	51		
Particles >21µm		ASTM D7647		9		
Particles >38µm		ASTM D7647	>10	0		
Particles >71µm		ASTM D7647		0		
Oil Cleanliness		ISO 4406 (c)	>19/17/14	17/16/13		
FLUID DEGRADA	TION	method	limit/base	current	history1	history2
Acid Number (AN)	mg KOH/g	ASTM D8045	1.10	0.66		


Report Id: LEPNEW [WUSCAR] 06197614 (Generated: 06/05/2024 16:20:15) Rev: 1


Submitted By: VINCENT MCINTIRE




# **OIL ANALYSIS REPORT**

VISUAL







		*Visual	NONE	NONE		
		*Visual	NONE	NONE		
		*Visual	NONE	NONE		
		*Visual	NONE	NONE		
		*Visual	NONE	NONE		
		*Visual	NONE	NONE		
		*Visual	NORML	NORML		
	scalar *	*Visual	NORML	NORML		
Emulsified Water	scalar *	*Visual		NEG		
Free Water	scalar *	*Visual		NEG		
FLUID PROPERTI	ES	method	limit/base	current	history1	history
Visc @ 40°C	cSt /	ASTM D445	220	221		
SAMPLE IMAGES		method	limit/base	current	history1	history
Color					no image	no image
Bottom					no image	no image
GRAPHS						
Ferrous Alloys			491,520	Particle Count		
iron						
E 6+ mickel			122,880			
			30,720	Devele		
2						
				Abnormal		
May28/24			May28/24 s (per 1 ml			
—			Ma cles (i		Ť	
Non-ferrous Metals			42/82/em Manuher 170			
8 copper			ja 120			
E 6+			E 30			
			30			
2						-
0						
May28/24			May28/24			
			2 2 0	4μ 6μ	14µ 21µ	38µ 71
Viscosity @ 40°C				Acid Number	- <i>p</i>	
Abnormal			( <sup>2.50</sup>	Abnormal		
240 + +			<u>ğ</u> 2.00			
() 0(+) 1220 - Base 133			 ອ	Base		
# I			(6)H00 2.50 bit 1.50 bit 1.50 Virginia 1.00 Virginia 1.00	1		
200 -				Abnormal		
200 - Abnormal			U.UU			
Abnormal			3/24	22		
Abnormal			/lay28/24	/lay28//		
200 Abnormal 180 + 200 4 200 4 200 4 200 4 200 4 200 4 200 4 200 4 200 4 200 4 200 200			May28/24	May28/24		
: WearCheck USA - 501			, NC 27513			
2001 200 200	Receive	ed : 03	7, NC 27513 3 Jun 2024		560	00 OMAHA
: WearCheck USA - 501 : WC0859498 : 06197614	Receive Tested	ed : 03 : 04	, NC 27513 3 Jun 2024 4 Jun 2024	I	560	00 OMAHA ROSWELL, I
2001 1801 1801 2006 2007 200 200	Receive Tested Diagno	ed : 03 : 04 osed : 05	7, NC 27513 3 Jun 2024	I ug Bogart	560	00 OMAHA ROSWELL, I US 882

To discuss this sample report, contact Customer Service at 1-800-237-1369.

\* - Denotes test methods that are outside of the ISO 17025 scope of accreditation.

Statements of conformity to specifications are based on the simple acceptance decision rule (JCGM 106:2012)

Report Id: LEPNEW [WUSCAR] 06197614 (Generated: 06/05/2024 16:20:16) Rev: 1

Certificate L2367

Submitted By: VINCENT MCINTIRE

Page 2 of 2

T: (575)347-3261

F: (505)347-5728